



Government of Nepal
Ministry of Home Affairs
Nepal A.P.F Hospital, Procurement Section,
Balambu, Kathmandu, Nepal

Amendment Notice

Date: 2078/08/26

(Dr. Rupan Maharjan)
Deputy Inspector General of
Armed Police Force, Nepal
Hospital Chief

1. Reference to our notice (IFB No: APFH/NCB/G/04/2079/080) Dated 2079/08/08 published on Nepal Samachar Patra Daily Newspaper. This is to inform all concerned parties that following point of Technical Specification has been amended.

Changes made in General Surgery Equipment (Endo-Urology Set up) (Contract ID. No. APFH/NCB/G/04/002/2079/080)		
S.No	Current Specification	Amended Specification
4.1	3CCD chip (3x1/2"), Digital Camera-PAL	3 CCD chip (3x1/2") or CMOS Digital camera
	Digital and Optical Zoom function which can be activated via both from camera head as well as from control unit .	Digital and Optical/Analog Zoom function which can be activated via both from camera head as well as from control unit .
	2 pre-set or programmable function keys on the camera head to control camera functions,	2 or more pre-set programmable function keys on the camera head to control camera functions.
	The camera should be quoted with the CCU, camera head, connecting cables for printers and recorders,	The camera should be quoted with the CC U with inbuilt USB HD video and image recording system, camera head, connecting cables for printers and recorders,

All other terms and conditions of above mentioned IFB No. APFH/NCB/G/04/002/2079/080 shall remain the same.