

Government of Nepal
Ministry of Home Affairs
Nepal A.P.F Hospital, Procurement Section,
Balambu, Kathmandu, Nepal

Amendment Notice

Date: 2080/08/10



1. Reference to our notice (IFB No: APFH/NCB/G/03/2080/081) Dated 2080/07/15 published on Nepal Samachar Patra Daily Newspaper. This is to inform all concerned parties that following point of Technical Specification has been amended.

| Changes made in Radiology equipments (Five Hundred MA x-ray Machine & DR System) (Contract ID. No. APFH/NCB/G/03/2080/081/005) | | |
|---|--------------------------------|--------------------------------|
| Five Hundred MA x-ray Machine | | |
| S.No | Current Specification | Amended Specification |
| 2.4 | Adone speed : 2800 rpm or more | Adone speed : 2700 rpm or more |

| DR System | | |
|------------------|--|--|
| S.No | Current Specification | Amended Specification |
| 3.1 | Flat panel detector system with computer and dry laser printer- 1 unit | Flat panel detector system with computer and dry image printer- 1 unit |
| 4.5 | Printer Should have dry Laser imager technology | Printer should have dry Laser/ Thermal Imager technology |

Note:-

Fixed Price list of necessary consumables for five years should be provided along with letter of price bid. Submitted price list of necessary consumables will be invalid in technical bid.

All other terms and conditions of above mentioned IFB No. APFH/NCB/G/03/2080/081 shall remain the same.