


Government of Nepal
Ministry of Home Affairs
Nepal A.P.F Hospital, Procurement Section,
Balambu, Kathmandu, Nepal
Amendment Notice
Date:2079/10/10



1. Reference to our notice (IFB No: APFH/NCB/G/05/2079/080) Dated 2079/09/24 published on Nepal Samachar Patra Daily Newspaper. This is informed to all concerned parties that the following points of Technical Specification in bid document has been amended.

Changes made High End Video Endoscopy System with Gastroscope and Colonoscope (Contract ID. No. APFH/NCB/G/05/2079/080)		
Tender Point No	Technical specification as published in tender	Amended specifications
4.12	Multi light LED (at least 4 LEDs) light technology shall be available	Multi light LED (at least 4 LEDs) or multi light xenon light technology shall be available. The supplier should give comprehensive warranty of the proposed light technology for minimum 7 years and should replace the light source as soon as its intensity deteriorates.

All other terms and conditions of above mentioned **Contract ID. No. APFH/NCB/G/05/2079/080** shall remain the same.