



Nepal A.P.F Hospital, Procurement Section,  
Balambu, Kathmandu, Nepal

**Amendment Notice**

Date: 2079/01/12

Reference to our notice (IFB No: APFH/NCB/G/04/2078/079) dated 2078/12/22 published on Rajdhani Daily Newspaper. This is to inform all concerned parties that some points of technical specification in bid document has been amended.

For detail amended notice please visit [www.bolpatra.gov.np/egp](http://www.bolpatra.gov.np/egp) or [www.apf.gov.np](http://www.apf.gov.np). or visit Nepal APF Hospital, Procurement Section, Balambu Kathmandu or contact Tel 01-4315224.

All other terms and conditions of the IFB No. APFH/NCB/G/03/2078/079 shall remain the same.

Government of Nepal  
Ministry of Home Affairs  
**Nepal A.P.F Hospital, Procurement Section,**  
Balambu, Kathmandu, Nepal  
**Amendment Notice**  
Date: 2079/01/12



1. Reference to our notice (IFB No: APFH/NCB/G/04/2078/079) Dated 2078/12/22 published on Rajdhani Daily Newspaper. This is to inform all concerned parties that following points of Technical Specification in bid document has been amended.

<b>Changes made Cardiology Equipments (TMT &amp; Holter ECG Machine)</b> (Contract ID. No. APFH/NCB/G/04/001/2078/079)		
<b>TMT Machine</b>		
Tender point no.	Technical specifications as published in tender	Amended specifications
4.6	System should preferably have options to use both wire and wireless (wifi based) stress test	System should mandatorily have wireless stress test system and should be compatible for wired stress test for future use
4.7	It should have real time ECG transmission via WIFI	It should have real time ECG transmission via wireless technology (Wifi or Bluetooth)
4.9	Full set of TMT i.e. software, accumulators, treadmill, etc. should be of same company	Full set of TMT i.e. software, accumulators, treadmill, automatic BP recorder should be of same company
4.24 c	Have elevation range 0-40 degree (please specify)	Have elevation range 0-25 degree (please specify)
7.2	European CE (93/42 EEC directives) and USFDA approved product certificate. Relevant certificate should be attached for verification	European CE (93/42 EEC directives) AND/OR USFDA approved product certificate. Relevant certificate should be attached for verification
<b>Holter ECG Machine</b>		
Tender point no.	Technical specifications as published in tender	Amended specifications
2.1	Holter monitor with 3, 7 and 12 channel digital recorder with removable SD card as well as USB transfer and Bluetooth	Holter monitor with 3 and 12 channel digital recorder (derived 12 channel digital recorder will be accepted) with removable SD card as well as USB transfer and Bluetooth.
4.2	3, 7 and 12 channel arrhythmia analysis	3 and 12 channel arrhythmia analysis, derived 12 channel arrhythmia analysis will be accepted.
4.11	It should be water resistance IPX4 or better	It should be water resistance IP22 or better. Holter machine should have water protection facility. The higher the protection, the better and will be prioritized
4.16	It should have function to take a snap shot of a trend graph and add it to report	It should have feature of adding event picture to the report
4.17	Holter point no. 4.17 patient protection against electric shock hazard from USB cable > 3000 V. Holter system operates in not more than 4 volt with DC line. So this point should be omitted	This point has been removed.
4.23	Digital recorder must have a minimum of 512 samples/sec/channel for pacemaker spike detection and at least 128 samples/sec/channel for recording and storage of 1000/sec/channel for VLP	Digital recorder must have 512 samples/sec/channel <b>or more</b> for pacemaker spike detection and at least 128 samples/sec/channel <b>or more</b> for recording and storage of 1000/sec/channel <b>or more</b> for VLP.

All other terms and conditions of above mentioned **Contract ID. No. APFH/NCB/G/04/001/2078/079** shall remain the same.

Government of Nepal  
Ministry of Home Affairs  
Nepal A.P.F Hospital, Procurement Section,  
Bafanbu, Kathmandu, Nepal  
**Amendment Notice**  
Date: 2079/01/12



2. Reference to our notice (IFB No: APFH/NCB/G/04/2078/079) Dated 2078/12/22 published on Rajdhani Daily Newspaper. This is to inform all concerned parties that following points of Technical Specification in bid document has been amended.

<b>Changes made in Neurodrill with craniotome (Contract ID. No. APFH/NCB/G/04/002/2078/079)</b>		
Tender point no.	Technical specifications as published in tender	Amended specifications
3.1	Must be light weight below 150g	Must be light weight below 200g
3.2	The high-speed motor should be powerful with power of around 150W and torque around 40N-m and should be in design to reduce vibration. Supporting document from principal company to prove the torque should be attached	The high-speed motor should be powerful with power of around 150W-200W and torque around 4.0N-m and should be in design to reduce vibration. Supporting document from principal company to prove the torque should be attached

All other terms and conditions of above mentioned **Contract ID. No. APFH/NCB/G/04/002/2078/079** shall remain the same.

Government of Nepal  
Ministry of Home Affairs  
Nepal A.P.F Hospital, Procurement Section,  
Balambu, Kathmandu, Nepal  
Amendment Notice  
Date: 2079/01/12



1. Reference to our notice (IFB No: APFH/NCB/G/04/2078/079) Dated 2078/12/22 published on Rajdhani Daily Newspaper. This is to inform all concerned parties that following points of Technical Specification in bid document has been amended.

Changes made in Ortho Drill and Saw Set (Contract ID. No. APFH/NCB/G/04/003/2078/079)		
Tender point no.	Technical specifications as published in tender	Amended specifications
1.1	Cutting speed should be at least 23000 CPM	Cutting speed should be at least 20000 +/- 5000 CPM
1.1	Reaming speed should be at least 330 RPM	Reaming speed should be at least 300 RPM
1.1	Drill speed should be at least 1330 rpm	Drill speed should be at least 1300 RPM
1.1	The diameter of cannulation should be at least 3.2 mm	The diameter of cannulation should be at least 3.5mm +/- 0.5mm
1.1	Weight of hand piece should not be more than 600-650 gm	Weight of hand piece should not be more than 600-650 gm without battery
	Added point	Commitment letter from distributor for availability of backup unit within 12 hrs. of break down

All other terms and conditions of above mentioned **Contract ID. No. APFH/NCB/G/04/003/2078/079** shall remain the same.